


How to Upgrade the AWA with Windows XP Embedded

Note: Also this version of the database is not compatible with Version 4.4.1 or older database. It is recommended that all instances of the software be upgraded to this version (4.4.7) of the software or databases cannot be shared. You will need a different install for the Desktop version of the AWA. This upgrade is for AWA's with Windows XP Embedded only.

Two modes are available on the AWA XP Embedded instrument– the Baker admin (has start menu and task bar) and User shell (AWA application is the shell, no menu nor taskbar). In order to install the AWA upgrade the instrument must be in the Admin shell. To switch to the Admin shell from the User Shell hold down the left **Alt** key and press **F12** while in the AWA application. The following dialog will appear, select the Admin mode and reboot.

The AWA will boot into the Explorer shell with a start menu and task bar. Connect the USB stick with the awaxpeupgrade.exe on it. Open the drive in Windows Explorer and double click the



awaxpeBetaUpgrad.exe, this will initiate the upgrade. By running the awaxpeBetaUpgrade.exe an AWA Update dialog will appear with a progress bar.



When the upgrade is complete the following dialog will appear indicating that the upgrade was completed.



Once the upgrade is completed the AWA will need to be switched back to the User Shell. To switch back to the User Shell, double click on the Baker_Admin short cut on the desktop. This will activate the Baker Admin dialog, click on the User Shell and Reboot.